

# EU Type Examination Certificate

This is to certify that:

Dromex  
Unit 1, 1 Blase Road  
New Germany  
3620  
South Africa

Holds Certificate Number:

CE 725172

In respect of:

**Disposable face masks to EN 149:2001+A1:2009  
Various models - see continuation sheets for details**

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2020-05-20

Latest Issue: 2020-05-20

Drs. Dave Hagenaaars, Managing Director

Effective Date: 2020-05-20

Expiry Date: 2025-05-20

Page: 1 of 4



...making excellence a habit.™

# EU Type Examination Certificate

No. CE 725172

## Product Specification

<b>Description:</b>	<b>Moulded Products with Option D clogging</b>
<b>Technical specification:</b>	EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles. Classification – Single shift NR
<b>FFP1 Models:</b>	1010
<b>FFP2 Models:</b>	1020 1021 (with exhalation valve fitted) 1121 (with activated carbon and exhalation valve fitted) 1121A (with activated carbon and exhalation valve fitted) 1121C (with activated carbon and exhalation valve fitted)
<b>FFP3 Models:</b>	1131 (with activated carbon and exhalation valve fitted) 1231 (with exhalation valve fitted)
<b>Description:</b>	<b>Vertical Fold Flat Products with Option D clogging</b>
<b>Technical specification:</b>	EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles. Classification – Single shift NR
<b>FFP2 Models:</b>	3020 3021 (with exhalation valve fitted)
<b>FFP3 Models:</b>	3231 (with exhalation valve fitted)

First Issued: 2020-05-20  
Latest Issue: 2020-05-20

Effective Date: 2020-05-20  
Expiry Date: 2025-05-20

Page: 2 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.  
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands  
A member of BSI Group of Companies.

# EU Type Examination Certificate

No. CE 725172

## Product Specification (continued)

<b>Description:</b>	<b>Vertical Fold Flat Products with Option D clogging</b>
<b>Technical specification:</b>	EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles. Models come with welded elastic straps. Classification – Single shift NR
<b>FFP2 Models:</b>	4020 4021 (with exhalation valve fitted)



First Issued: 2020-05-20  
Latest Issue: 2020-05-20

Effective Date: 2020-05-20  
Expiry Date: 2025-05-20

Page: 3 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.  
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands  
A member of BSI Group of Companies.

# EU Type Examination Certificate

No. CE 725172

## Certificate Administration Details

Product initially approved on BSI certificate: CE 690046

## Certificate Amendment Record

Issue Date	Comments
------------	----------

May 2020	First Issue
----------	-------------

**BSI Internal  
Project Number**

2792:20:3153912

## Certificate validity

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 60966.

First Issued: 2020-05-20

Latest Issue: 2020-05-20

Effective Date: 2020-05-20

Expiry Date: 2025-05-20

Page: 4 of 4

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.  
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands  
A member of BSI Group of Companies.